

Specifications of different grades of stainless steel used in HPLC columns

Cogent™ HPLC stainless steel column hardware is T1 grade, AISI 316 LC type. HPLC system tubing is AISI 316 (highlighted in green in the chart below, T2 grade. *We also offer fully coated stainless steel column hardware and PEEK column hardware to eliminate sample interaction with the surface sites of stainless steel.*

Elemental assays of the various stainless steel grades are given in the following table:



Ti Grade 2

Carbon	0.1 max
Hydrogen	0.015 max
Iron	0.3 max
Nitrogen	0.03 max
Oxygen	0.25 max
Titanium	Balance



MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com