

Establishing HPLC method linearity – Tips & Suggestions

To establish linearity of the proposed method, ten point calibration curves should be generated with appropriate concentrations of calibration standard solutions. The calibration range can be for example: 10-100 microgram / mL.

The linearity should be evaluated by the least squares regression method using unweight data.