



Can amber or clear polypropylene vials be sterilized with autoclaving or ethylene oxide - FAQ

Polypropylene autosampler vials [9502S-PP-Clear](#) should not be autoclaved because they may warp during heating. However, you can use ethylene oxide gas to sterilize the vials if you do not exceed 120°C.

Click [HERE](#) for more information about polypropylene vials and pictures

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024