

Can amber or clear polypropylene vials be sterilized with autoclaving or ethylene oxide – FAQ

Polypropylene autosampler vials [9502S-PP-Clear](#) should not be autoclaved because they may warp during heating. However, you can use ethylene oxide gas to sterilize the vials if you do not exceed 120°C.

Click [HERE](#) for more information about polypropylene vials and pictures