

If you observe changing retention times of amino acids using a Cogent Diamond Hydride™ HPLC column, consider the following.**Suggestions, tips and tricks**

1. Acid in the Mobile Phase; Formic or acetic acid is generally used when analyzing amino acids, so the use of ammonium acetate (or formate) might be a problem.
2. Contamination from the biological samples. Try our conditioning protocol for biological samples.
3. Suggested Mobile phase: 50% methanol or 50% isopropanol +0.1 % formic acid or 0.2% acetic acid would be suggested to use with biological samples and amino acid analysis when using the Cogent TYPE-C™ Silica columns.
4. How long is your gradient's re-equilibration time? Follow our protocol on choosing a proper post time.
5. Is the column thermostated? If not, the variation may be due to changes in lab temperature.

[Cogent Diamond Hydride HPLC Column Product Page](#)



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com