
MICROS

Carotenoids Analyzed with HPLC - AppNote

Compounds of the Carotenoid Family Retained

These compounds are very lipophilic and often have subtle differences in structure. Resolution of these compounds can be difficult with a standard Reversed Phase C8 or C18 Column because of these Similarities. The Cogent C30 Column on the other hand can further differentiate by analyte Shape in addition to Reversed Phase Interactions. At lower temperatures, the long alkyl chains become more rigid and steric effects become significant, leading to greater Selectivity.



- 13. d-Carotene
- 14. Lycopene

Method Conditions:



Column: Cogent C30[™], 5µm, 200Å

Catalog No.: 71030-25P

Dimensions: 4.6 x 250mm

Mobile Phase:

A: 81:15:4 Methanol / Methyl Tert-Butyl Ether / DI Water (v/v) B: 6:90:4 Methanol / Methyl Tert-Butyl Ether / DI Water (v/v)

Gradient:

Time (minutes)	%B
0	0
90	100

Temperature: 20°C Flow rate: 1.0ml / minute Detection: UV @ 450nm Sample Preparation: Reference Standards of each Analyte in a mixture.

Notes: Carotenoids are a broad class of more than 600 compounds. They can be divided into two types: Xanthophylls and Carotenes. Xanthophylls contain oxygen in their structure and get their name from the Greek words Xanthos (yellow) and Phyllon (leaf). Carotenes are unsaturated hydrocarbons which do not contain other elements in their structure.



Attachment

No 262 Carotenoids Analyzed with HPLC pdf 0.2 Mb Download File

Printed from the Chrom Resource Center **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435 Email: customers@mtc-usa.com Website: www.mtc-usa.com Date: 06-05-2024