

## Can you inject highly acidic samples on to Cogent TYPE-C Silica HPLC columns - FAQ

*The Cogent TYPE-C™ Silica columns have certain specification ranges for pH that must be followed to avoid damage to the material. These specifications apply to the mobile phase in the column.*

The amount of sample injected is so small compared to the mobile phase that the effect is not usually significant. Therefore you should be able to inject highly acidic samples without decreasing the column's lifetime as long as the mobile phase is within column pH specifications.

However, you may observe peak shape distortions if the mobile phase pH is much different than the sample pH. This does not mean the column was damaged but can happen any time the [sample diluent and mobile phase pH differ](#) significantly.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024