



What is the maximum pressure of the 2.1mm PEEK HPLC columns - FAQ

The maximum recommended pressure of the PEEK HPLC column under routine use is 165 bar (2400 psi).

The MicroSolv 2.1mm ID PEEK column is made of 1/8" tubing. However, when packing the stationary phase, a pressure of up to ~250 bar (3600 psi) can be applied for a short time.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com