

The capillaries should be stored at room temperature (approximately 22°C) in a dry place. The temperature these capillaries are stored at does not depend on the ID of the capillary. Many customers actually tape them to the wall near their **CE** instrument.



Click [HERE](#) for Zero Flow capillaries ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)