

Can I use a laser to make a detection window for the Controlled Flow and Zero Flow capillaries – FAQ

No, these capillaries have a special, covalently bonded coating on the inside capillary wall (linear polyacrylamide) that may be damaged using a laser or heat.

The outer polyimide coating can be removed by a *specific protocol for this purpose*. This procedure uses fuming sulfuric acid to chemically remove the outer polyimide coating.



Click [HERE](#) for ordering information and pictures.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com