

Effect of diluent on analyte adsorption to glass autosampler vials – White Paper

It is well known that silanol groups present on the surface of Type I glass autosampler vials contribute to quantitative loss in the case of basic analytes. This is believed to occur primarily by electrostatic interaction and possibly by hydrogen bonding to some degree.

In order to investigate the nature of this interaction in more detail, we conducted a series of experiments using different sample diluents and used RSA™ Reduced Surface Activity autosampler vials.

Download the attached PDF to read full article using the link below.



Click [HERE](#) for RSA vials ordering information and pictures

Attachment

Effect of diluent on adsorption pdf [Download File](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-14-2024