



Avoid loss of concentration of heparin with HPLC when using syringe filters – Tips & Suggestions

The most appropriate membrane type to use when filtering heparin containing samples is AQ™ brand regenerated cellulose (RC) syringe filter.

Heparin has strongly acidic sulfonate groups (*negatively charged*) that will bind with nylon filter membranes. PTFE is not suitable because of its hydrophobicity and passes water through it is not possible.

Click [HERE](#) for ordering information and pictures of AQ brand of syringe filters.

 25mm Regenerated Cellulose Filter

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com