


## Benzylpiperazine (BZP) – AppNote


### Analysis in Hair Samples using LCMS

An analytical protocol was developed for analysis of Benzylpiperazine in human hair samples. The elucidation of the chromatographic method was challenging due to the polar nature of BZP. The limits of detection / quantification for this method were determined to be 0.05 ng / mg for Benzylpiperazine in hair samples.

The method was found to be linear from 0.1 – 10 ng/mg ( $r^2 > 0.999$ ). Recovery of Benzylpiperazine was found to be greater than 95%. Matrix effects were determined to be < 6%.

The concentration of Benzylpiperazine in spiked samples of hair was determined in range from 1.2 – 1.5 ng/mg. The procedure after validation will be useful for laboratories performing routine analysis of drugs of abuse.





Benzylpiperazine

## Peaks:

A: Benzylpiperazine 177.1386 m/z [M+H]<sup>+</sup>

B: 5 injections of the sample

## Method Conditions

**Column:** Cogent Diamond Hydride™, 4μm, 100Å

**Catalog No.:** 70000-15P-2

**Dimensions:** 2.1 x 150mm

### Mobile Phase:

A: DI Water / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

### Gradient:

Time (minutes)	%B
0	80
3	30
6	30
10	80

**Flow rate:** 0.4 mL/minute

**Detection:** ESI - POS - Agilent 6210 MSD TOF Mass Spectrometer

**Injection vol.:** 1μL

**Sample Preparation:** Extraction (SPE) was performed on cartridge I (Clean Screen Xcel™ purchased from UCT Bristol, PA, USA), preconditioned with 3mL of Methanol, 3 mL of DI Water, and 1 mL of pH 6 buffer prior to sample loading. 10 mg samples of hair (controls, and spiked test samples) were digested in 1 mL of 1 M Sodium Hydroxide for 1 hour at 70°C. The samples were cooled, and 100 μL of Acetic Acid (Glacial) was added.

Each solution was adjusted to pH 6 with 0.1 mM Ammonia and applied to the SPE column. After loading the samples, each sorbent was washed with DI Water, Acetic Acid (0.1 M), and Methanol (3 mL of each, respectively). Each SPE Column was dried and eluted with 3 mL of Methylene Chloride / Isopropanol / Ammonium Hydroxide (78:20:2). After elution, solvents were evaporated and 200 μL of Mobile Phase was added. The samples were used for analysis by LCMS.



**Attachment**

**No 321 Benzylpiperazine pdf** 0.2 Mb [Download File](#)

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-08-2024