
Are septa in autosampler vial caps self sealing, inert and compatible with organic solvents?

All of our septa are self-sealing and lined with PTFE or other fluoropolymer for inertness. These solvents do not present a compatibility issue but acetonitrile and ethers are very volatile and may evaporate after a single use or piercing of the septa. You may need to replace the cap after an injection if you need extreme control over evaporation from the vial.

[Click HERE for Screw Cap Ordering Information including Soft-Guard](#)