



To obtain greater sensitivity than Na adducts for sugars could you add ammonia NH₄ instead - FAQ

In LCMS, you will get much better sensitivity of Na-sugar adducts in the EICs compared to sugars themselves.

The same principle will work for NH₄ adducts as well, but there are notable drawbacks. The problems with NH₄ is that it will change the pH of the mobile phase and its concentration is hard to control in the sample. Hence, given both strategies will work, there is no reason to choose ammonia NH₄ over sodium Na.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-17-2024