



Is the HSQ Kit compliant with the latest USP-NF chapter revisions for 857 and 1857 - FAQ

The MicroSolv HSQ™ Kit materials are NIST traceable.

Compliance with USP protocols is demonstrated by the user and not the kits themselves. Specific chapters in the USP refer to procedures and the HSQ™ Kit materials can be used as part of your protocol to meet the relevant USP requirements.

For example, the MicroSolv UV and HPLC performance qualification kits can be used to comply with USP <1058> or USP <857> regarding the qualification of instruments.

Click [HERE](#) for PQ & HSQ ordering information and pictures of the products.