



Have the amber autosampler vials changed color compared to earlier batches – FAQ

The amber vials catalog # 9502S-PP-AMBER have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # [9502S-PP-AM](#). You may notice the newer vials are more opaque.

[Click HERE for 9502S-PP-AM Ordering Information](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-18-2024