



## Have the amber autosampler vials changed color compared to earlier batches - FAQ

The amber vials catalog # *9502S-PP-AMBER* have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # *9502S-PP-AM*. You may notice the newer vials are more opaque.

*Click [HERE](#) for 9502S-PP-AM Ordering Information*

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024