



Have the amber autosampler vials changed color compared to earlier batches - FAQ

The amber vials catalog # *9502S-PP-AMBER* have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # *9502S-PP-AM*. You may notice the newer vials are more opaque.

Click [HERE](#) for 9502S-PP-AM Ordering Information

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024