



Have the amber autosampler vials changed color compared to earlier batches - FAQ

The amber vials catalog # *9502S-PP-AMBER* have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # *9502S-PP-AM*. You may notice the newer vials are more opaque.

[Click HERE for 9502S-PP-AM Ordering Information](#)