

Have the amber autosampler vials changed color compared to earlier batches - FAQ

The amber vials catalog # 9502S-PP-AMBER have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # 9502S-PP-AM. You may notice the newer vials are more opaque.

Click HERE for 9502S-PP-AM Ordering Information

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com Website: www.mtc-usa.com