



Aegeline in Protein Sport Mix – AppNote

LC-MS analysis of harmful substance in food supplement

The data demonstrates the use of the Cogent Bidentate C18 2.0™ Column for analysis of dietary supplement samples. Isolation of the Aegeline peak is readily obtained with an extracted ion chromatogram (EIC) in the positive ion mode. Chromatographically, the peak is well retained and highly symmetrical. The Column is an excellent choice for analysis of this controversial compound in dietary supplements such as these.



Aegeline

Peak:

Aegeline 298.1438 m/z [M + H]⁺

Method Conditions

Column: Cogent Bidentate C18 2.0™, 2.2µm, 120Å

Catalog No.: 40218-05P-2

Dimensions: 2.1 x 50 mm

Mobile Phase:

A: DI Water / 0.1% Formic Acid

B: Acetonitrile / 0.1% Formic Acid

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Time (minutes)	%B
0	20
6	90
8	90
10	20

Injection vol.: 1 μ L

Flow rate: 0.4mL/minute

Sample: Two scoops (2.9456g) of a protein powder mix were dissolved in 50.00 mL DI Water in beaker. A portion was filtered (0.45 μ m, Nylon) and 400 μ L filtrate was mixed with 400 μ L Acetonitrile. The solution was centrifuged and the supernatant was collected and used for LC-MS injections.

Detection: ESI - POS - Agilent 6210 MSD TOF Mass Spectrometer

Note: *Aegeline is an alkaloid extracted from Aegle marmelos (bael), a plant that has been used in traditional medicines. It is also synthesized and has been added to some food supplements such as OxyElite Pro®. Although it is claimed to have weight-loss effects, its safety has not been thoroughly evaluated in humans and has even been implicated in instances of acute liver damage and non-viral hepatitis.*



Attachment

No 336 Aegeline in Protein Sport Mix pdf 0.4 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com