

What is the inside diameter of the opening of your Luer Lock syringes catalog number 58901-S - FAQ

The inner diameter (ID) of the opening of the Chrom Syringes has been measured to be approximately 2.1-2.2mm.



[Syringes for Chromatography Ordering Information](#)