
What is the Inside diameter of the Opening of your Luer-Lock Syringes Catalog# 58901-S?

The Inner Diameter (ID) of the opening of the Chrom Syringes has been measured to be approximately 2.1-2.2mm.



[Syringes for Chromatography Ordering Information](#)