



## What type of tubing do I use to connect the guard column holder to analytical columns - FAQ

**The best choice is to use a column coupler to minimize dead volume and not tubing.** As for the appropriate ID, you should use the same ID as in your HPLC system tubing if this is how you want to proceed.

Click [HERE](#) for column couplers ordering information and pictures.