



## What type of tubing do I use to connect the guard column holder to analytical columns - FAQ

**The best choice is to use a column coupler to minimize dead volume and not tubing.** As for the appropriate ID, you should use the same ID as in your HPLC system tubing if this is how you want to proceed.

Click [HERE](#) for column couplers ordering information and pictures.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-27-2024