
What type of tubing do I use to connect the guard column holder to analytical columns? - FAQ

The best choice is to use a column coupler to minimize dead volume and not tubing. As for the appropriate ID, you should use the same ID as in your HPLC system tubing if this is how you want to proceed.

Click [HERE](#) for Column Couplers Ordering Information

