

## Buffer is not eluting out of the Zero Flow capillary is the capillary defective – FAQ

Zero Flow™ Capillaries supplied by MicroSolv are designed to be Electro-Osmotic-Flow (EOF)-free. This is due to a covalently bonded inner coating that eliminates surface hydroxyl group **activity** responsible for bulk flow or EOF. Therefore it is normal to observe “no **eluent**” of **buffer** or any liquid even at high voltages. This can be beneficial because uncontrolled EOF is known to contribute to a lack of **robustness** in **CE**.

NOTE: When **conditioning** the capillary, pressure may be used to push **buffer** through the capillary. In this case, you will see **buffer** coming out. If not, then there is likely a blockage in the capillary which is an issue so please contact our customer service department at customers@mtc-usa.com or call us 1-732-380-8900 between 9:00AM and 5:30PM EST.

Movement of the Cationic or Anionic analytes through the capillary from inlet to electrode is due to their electro mobility only and EOF is not involved.



Click [HERE](#) for ordering information and pictures

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024