



How to improve peak shapes, retention & carryover for polar compounds with the Cogent Diamond Hydride™ column – Tips & Suggestions

Peak shape for polar compounds analysis such as vitamins in LCMS has been shown to be improved when using a 0.1% formic acid additive instead of ammonium acetate.

The sample diluent should be 50% DI water/ 50% acetonitrile/ 0.1% formic acid. If you are using 50% isopropanol or 50% methanol in the aqueous solvent, the carryover could be due to insufficient eluting strength.

Try using DI water/ 0.1% formic acid instead of 50% 2-propanol/ 50% DI water/ 0.1% formic acid.

[Cogent Diamond Hydride HPLC Column Ordering Information](#)



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-14-2024