



## How do I clean the PQ or HSQ kit HPLC column - FAQ

Since the column should only be used for test purposes, it should never become contaminated, and should require only a minimal 10-15 minute equilibration prior to starting the sequence. If it ever requires cleaning, flush for 15-30 minutes with 50:50 acetonitrile : water.”

Click [HERE](#) for PQ & HSQ Kit ordering information and pictures.