
How do I clean the PQ or HSQ Kit™ Column?

“Since the column is used only for test purposes, it should never become contaminated, and should require only a minimal 10-15 minute equilibration prior to starting the sequence. If it ever requires cleaning, flush for 15-30 minutes with 50:50 acetonitrile : water.”

—User manual, Rev. 1.08, section II.3, pg. 9.

[PQ & HSQ Kits for HPLC Ordering Information](#)