



## What do WCS and WVS stand for in Partisil HPLC columns - FAQ

WCS = Whatman Compression Sealing hardware

WVS = Whatman Void Sealing hardware

*Void sealing technology allows for easy adjustment of the frit position to compensate for any voids that develop at the head of the column. This can prolong column life in the event that dissolution of a section of the packing bed occurs.*