



## Are the AQ brand of nylon syringe filters suitable for sterile filtration - FAQ

**The AQ™ syringe filters are not designed for sterile filtration.**

Their primary intended function is to remove particulate matter from a sample to be injected into and protect an HPLC instrument and the HPLC column.

Click [HERE](#) for ordering information and pictures of AQ™ brand syringe filters

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 08-31-2024