



Can I sterilize my sample by using a syringe filter – FAQ

When you use a 0.2 µm filter you remove any particle that is larger than 0.2 µm.

This includes bacteria as well as many viruses and spores. However, it is important to note that when you remove these interferences, and your sample passes through the membrane and enters the outlet of the filter device, your solution is not really sterile any longer, unless you have sterilized the device.

Using the 0.2µm filter is especially helpful for UHPLC methods as it is important to remove larger particles including bacterial that can clog the column frits which use very small pores.

Click [HERE](#) for ordering information and pictures of AQ brand of syringe filters.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-05-2024