



What type or brand of injector should I use with the 2 μ L sample loop - FAQ

The 2 μ L sample loop *catalog # 47001-002* can be used with **Rheodyne™ 9725 and 9725i** sample injectors which are suitable for a 2 μ L injection volume.

In addition, it is designed for use with a Rheodyne™ type injector and Not Valco™ types.

Click [HERE](#) for sample loop ordering information and pictures.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-18-2024