



## What type or brand of injector should I use with the 2 $\mu$ L sample loop - FAQ

The 2  $\mu$ L sample loop *catalog # 47001-002* can be used with **Rheodyne™ 9725 and 9725i** sample injectors which are suitable for a 2  $\mu$ L injection volume.

In addition, it is designed for use with a Rheodyne™ type injector and Not Valco™ types.

Click [HERE](#) for sample loop ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-27-2024