
What type or brand of injector should I use with the 2 μ L sample loop - FAQ

The 2 μ L sample loop *catalog # [47001-002](#)* can be used with Rheodyne™ 9725 and 9725i sample injectors which are suitable for a 2 μ L injection volume.

In addition, it is designed for use with a Rheodyne™ type injector and Not Valco™ types.

Click [HERE](#) for sample loop ordering information and pictures.