
Is the plastic used in the U-2D system autoclavable at 120°C - FAQ

The plastic material of the U-2D™ Micro-Sample Management system begins to melt around 135°C.

Since this is quite close to the typical autoclaving temperature of 120°C, we do not recommend or support the use of heat sterilization with this product. However, it is compatible with gas sterilization.

Click [HERE](#) for U-2D™ ordering information and pictures.