



## Is the plastic used in the U-2D system autoclavable at 120°C – FAQ

**The plastic material of the U-2D™ Micro-Sample Management system begins to melt around 135°C.**

Since this is quite close to the typical autoclaving temperature of 120°C, we do not recommend or support the use of heat sterilization with this product. However, it is compatible with gas sterilization.

Click [HERE](#) for U-2D™ ordering information and pictures.

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)