

How does ammonium fluoride affect chromatography using Cogent TYPE-C HPLC columns?

Ammonium fluoride as mobile phase additive functions beneficially in several ways in a for LCMS:

- As a buffer (*i.e.* pH control)
- Improvement of peak shapes
- Improvement of sensitivity in MS for certain compounds

For further reading, please refer to the following third-party journal article reference:

J. Pesek, M. Matyska, "Ammonium fluoride as a mobile phase additive in aqueous normal phase chromatography," J. Chromatogr. A. 1401, 2015, 69-74.

NOTE: care should be taken when using ammonium fluoride because it is corrosive to glass (Teflon bottles are preferred) and using a concentration in excess of 1 mM may permanently damage HPLC columns. However, it has been reported that Cogent TYPE-C™ Silica columns last much longer than ordinary columns but ammonium fluoride may somewhat decrease the column lifetime.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024