
What is the best way to attach the inlet of the knitted reactor coil KRC tubing to the outlet of a tee with a PEEK finger tight fitting that will hold high back pressure - FAQ

Confirm that there is not an impediment to reasonable flow through the tubing which could increase the back pressure. We would then recommend using our items [47320-30](#) and [47320-34](#) which are high pressure PEEK HPLC tubing connectors or male nuts and ferrules.