



Capillary storage in capillary electrophoresis – How To

When storing a capillary after use, follow these simple instructions to increase the capillary life.

1. Filter all solutions to be used on your capillary with .45 μ or .2 μ filter.
2. Rinse the capillary with .1M NaOH solution for 5 minutes.
3. Rinse the capillary with CEwater or CE grade water for 5 minutes.
4. If possible, blow nitrogen or helium through the capillary prior to storage.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-27-2024