

Capillary storage in capillary electrophoresis - How To

When storing a capillary after use, follow these simple instructions to increase the capillary life.

- 1. Filter all solutions to be used on your capillary with .45 μ or .2 μ filter.
- 2. Rinse the capillary with .1M NaOH solution for 5 minutes.
- 3. Rinse the capillary with CEwater or CE grade water for 5 minutes.
- 4.If possible, blow nitrogen or helium through the capillary prior to storage.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

> Email: customers@mtc-usa.com Website: www.mtc-usa.com