



## Effects of Changing EOF in Capillary Electrophoresis.

When you change the Electro Osmotic Flow (EOF) of your CZE method, the following changes can be expected.

**Joule Heating:** Reducing or lowering the EOF will reduce the amount of Joule Heating in your capillary.

**Migration Time:** Changes in EOF will directly effect the migration times of all of your analytes. The changes are not always proportional and this may affect your peak shapes and peak areas.

**Resolution:** When EOF is reduced or eliminated Resolution can be enhanced.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024