

## Troubleshooting Common Injection Problems in Capillary Electrophoresis

Some of the most common problems with HPCE is caused by problems or errors created during the injection procedure. Following is a short table of Problems and Causes and Solutions to these problems.

Problem	Cause/Solution
No Injection at all	<ol style="list-style-type: none"> <li>1. Replace the capillary or unplug the capillary</li> <li>2. Check the pressure or vacuum of the system</li> <li>3. Check to see if your sample vial is empty</li> <li>4. Check to see if you are injecting from the correct vial</li> <li>5. Check your polarity. Make sure it is set properly</li> <li>6. Call your instrument's service technician to check the injector</li> <li>7. The openings may be obstructed by residual coating material on a new capillary. Cut 1-2 mm off the ends in this case.</li> </ol>
Variable Current	<ol style="list-style-type: none"> <li>1. Large Injection causes this normally — reduce the amount you are injecting</li> <li>2. Plugged Capillary? It should be replaced</li> <li>3. Check to see if you are using the correct <b>buffer</b> vial</li> <li>4. You may have injected an air bubble. Recondition the capillary</li> <li>5. Check to make sure that the <b>analyte</b> and the catholyte are the same, at the same level and electrolysis did not take place due to <b>buffer</b> depletion.</li> <li>6. The openings may be obstructed by residual coating material on a new capillary. Cut 1-2 mm off the ends in this case.</li> </ol>
Poor Peak Area	<ol style="list-style-type: none"> <li>1. Inject more sample in subsequent runs</li> <li>2. Check to make sure that electrolysis did not occur due to <b>buffer</b> depletion. If so, replace the anolyte and the catholyte</li> <li>3. Check to make sure that there is sufficient sample in your injection vial. Watch out for evaporation as well</li> </ol>

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 04-05-2024