



## Does the method used in the HSQ Kit for HPLC need to be validated - FAQ

**Validation of the method would not be applicable since it is used to qualify the instrument itself.**

Suppose you're talking about LOD or LOQ of the HPLC method as a validation parameter, for example. You wouldn't need to validate this parameter in the HSQ Kit™ method because its calculation first assumes that the detector is responding in an accurate manner; however, detector **accuracy** is one of the parameters that the HSQ Kit™ itself qualifies. Validation of this HPLC method is not required by any federal regulation.

Please note that before you validate a method, you must complete IQOQPQ on the instrument.

Click [HERE](#) for PQ & HSQ Kits™ ordering information and pictures.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 07-22-2024