

What are the concentrations used in the Refractive Index qualification kit – FAQ

The solutions of the HPLC Performance qualification solutions for Refractive Index and Light Scattering Detectors have the following certified concentrations:

Solution	Header 2
L1	0.200 mg/mL \pm 0.0002
L2	0.400 mg/mL \pm 0.0002
L3	0.500 mg/mL \pm 0.0002
L4	1.00 mg/mL \pm 0.001
L5	1.50 mg/mL \pm 0.001
L6	2.00 mg/mL \pm 0.001

Calibration **Linearity** Standards: Caffeine dissolved in 10:90 acetonitrile:0.1% acetic acid.



Certificates of Analysis of specific lots can be furnished upon request (**contact [Customer Service](#)**)

Click [HERE](#) for qualification product ordering information and pictures.