



Does deactivation process of RSA-Pro X autosampler vials add any chlorides to the glass surface - FAQ

No, the RSA-Pro X™ surface treatments used do not add chlorides.

Click [HERE](#) for the RSA-Pro X technical information page with a demonstration video.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com