



## Does deactivation process of RSA-Pro X autosampler vials add any chlorides to the glass surface - FAQ

**No, the RSA-Pro X™ surface treatments used do not add chlorides.**

Click [HERE](#) for the RSA-Pro X technical information page with a demonstration video.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-05-2024