



# **Retained Isocratically or With a Gradient**

This cyclic Octapeptide can be retained with simple Isocratic Mobile Phase conditions of 50:50 Solvent A / Solvent B (see Figure I). In addition, use of a Gradient produces a sharper Peak if desired (Figure II). Note, that the Gradient starts at an unusually high percent of Water (50%) for an Aqueous Normal Phase (ANP) application, due to the highly polar nature of the molecule.

With its two basic Amino Acid residues, the Octapeptide may lead to tailing due to silanolic interactions on some conventional Silica-based HPLC stationary phases, but here the peak shape is very symmetrical.





Printed from the Chrom Resource Center Copyright 2024, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



#### Peak:

Octreotide

# **Method Conditions:**

**Column**: Cogent Diamond Hydride<sup>™</sup>, 4μm, 100Å

**Catalog No.:** 70000-15P-2 **Dimensions**: 2.1 x 150 mm

**Solvents**:

A: DI Water / 0.1% Formic Acid (v/v)
B: Acetonitrile / 0.1% Formic Acid (v/v)

Figure I: 50% solvent A/ 50% solvent B

## Figure II: Gradient:

Time (Minute)	%B
0	50
3	20
6	20
7	50

Injection Volume: 1µL Flow Rate: 0.4mL / minute

**Detection**: ESI - POS - Agilent 6210 MSD TOF Mass Spectrometer **Sample Preparation**: Reference standard solution of Octreotide

**Notes:** Octreotide mimics the naturally occurring hormone Somatostatin. It is used for the treatment of growth hormone producing tumors and other related medical applications.



### Attachment

No 367 Octreotide Analyzed with LCMS pdf 0.1 Mb Download File

Printed from the Chrom Resource Center
Copyright 2024, All Rights Apply
MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com